



Product Service

# Attestation of Conformity

No. E8A 076845 0007 Rev. 00

**Holder of Certificate:** **SENKO Co., Ltd.**  
73, Oesammi-ro, 15beon-gil  
Osan-si, Gyeonggi-do 18111  
REPUBLIC OF KOREA

**Name of Object:** **Gas Detection Systems  
(Portable Single Gas Detector)**

**Model(s):** **iGas Detector CO2**

**Description of Object:**

Rated input voltage:	100-240 V a.c. (adapter) 3.7 V d.c. (Battery)
Rated frequency:	50 / 60 Hz
Rated output voltage:	5 V d.c.

**Tested according to:**

EN 50270:2015  
EN 61000-6-4:2007/A1:2011  
EN 61000-3-2:2014  
EN 61000-3-3:2013

This Attestation of Conformity is issued on a voluntary basis according to the Directive 2014/30/EU relating to electromagnetic compatibility. It confirms that the listed apparatus complies with all essential requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for testing and certification. For details see: [www.tuvsud.com/ps-cert](http://www.tuvsud.com/ps-cert)

**Test report no.:** DREKCEE2005-0298

**Date,** 2020-05-22

( Byung-Soo Kang )

Page 1 of 1

After preparation of the necessary technical documentation as well as the EU Declaration of conformity the required CE marking can be affixed on the product. That Declaration of conformity is issued under the sole responsibility of the manufacturer. Other relevant EU-directives have to be observed.